



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/823,416	04/12/2004	Matthew R. Selmon	LUM 5004 USCNT3	3249

  

27777	7590	07/08/2010
PHILIP S. JOHNSON JOHNSON & JOHNSON ONE JOHNSON & JOHNSON PLAZA NEW BRUNSWICK, NJ 08933-7003		

  

EXAMINER	
GRAY, PHILLIP A	

  

ART UNIT	PAPER NUMBER
3767	

  

NOTIFICATION DATE	DELIVERY MODE
07/08/2010	ELECTRONIC

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

jnjuspatent@corus.jnj.com  
lhowd@its.jnj.com  
gsanche@its.jnj.com

<b>Office Action Summary</b>	<b>Application No.</b> 10/823,416	<b>Applicant(s)</b> SELMON ET AL.	
	<b>Examiner</b> Phillip Gray	<b>Art Unit</b> 3767	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 26 April 2010.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1,4,13,20,21,25 and 26 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1,4,13,20,21,25 and 26 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)                                | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftperson's Patent Drawing Review (PTO-948)                        | 5) <input type="checkbox"/> Notice of Informal Patent Application                       |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____  |

### **DETAILED ACTION**

This office action is in response to applicant's communication of 4/26/2010.

Elected claims 1, 4, 13, 20-21, 25-26 are pending and rejected below.

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 4, 13, 20-21, 25-26 are rejected under 35 U.S.C. 102(b) as being anticipated by Makower et al. (U.S. 6,190,353). Makower discloses a method and apparatus for bypassing arterial obstruction and performing other transvascular procedures with a system comprising a single guide wire (GW as in figure 10c, 10c", and 10c""") a catheter body with lumen (700), a catheter endpiece including a deflection housing (700 distal end as shown in figure 10c" and 10c"""), a least one distal port (distal most end of 700), one lateral opening (side port that element 102 and GW exits from in figure 10c") in communication with the lumen; a working element (curved sharpened distal tip near 728), note that the working element would be in a straight configuration when positioned in the catheter body prior to the working element leaving the lateral opening in a curved configuration, further note the deflection housing (near distal end of 700) would allow the guide wire to pass either the distal or lateral port (as shown in figure 2a). Concerning claim 4 it is examiners position that the catheter lumen 700 is

Art Unit: 3767

uniform. Concerning claim 13 it is examiners position that Makower working element distal tip near 728 contains a preformed resilient tip, see figure 10c" and 10c"".

Concerning claims 20-21 see figures 2A-5d, 6a-13e. Concerning claims 25-26 see column 46 lines 51-65 concerning handpiece 706 and 710.

### ***Response to Arguments***

Applicant's arguments filed 4/26/2010 have been fully considered but they are not persuasive. Applicant's argue that the amendments to the claims concerning the working element "...comprising a cannula with a lumen for traversing over the single guide wire and having a curved distla section and a sharpened distal tip..." and the curved distal section of the cannula being in a straight configuration when positioned within the catheter body and in a curved configuration when extending from the lateral opening, the deflection housing being configured to allow the guide wire to pass through either of the distal port or the lateral opening ..." is not shown in Makower. Examiner is of the position that Makower does disclose these limitations. See rejection above specifically. Examiner is of the position that the Makower catheter and working element would be in a straight configuration until the working element is deployed from the lateral distal opening and would leave the opening in a curved configuration. The elements disclosed in Makower are fully capable of satisfying all structural, functional, spatial, and operational limitations in the amended claims, as currently written, and the rejection is made and proper.

***Conclusion***

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Phillip Gray whose telephone number is (571)272-7180. The examiner can normally be reached on Monday through Friday, 8:30 a.m. to 4:30 p.m. EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Kevin Simons can be reached on (571) 272-4965. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3767

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Phillip Gray/  
Examiner, Art Unit 3767

/Kevin C. Sirmons/  
Supervisory Patent Examiner, Art Unit 3767